

ID Panel Profile



0843
IVD



Blood and Transplant

NHSBT REAGENTS Panel 2

Product	Lot No.	Product	Lot No.	Product	Lot No.
ID Panel in Alsevers	R141 3403	ID Panel Papainised in Alsevers	R15n/a	ID Panel in LISP	R14n/a
ID Panel in CellStab	R142 3403	ID Panel Papainised in CellStab	R152 3403	Expiry Date : 2018.06.14	
ID Panel in CellMedia	R16n/a	ID Panel Papainised in CellMedia	R17n/a		

Patient's Name	Ref. No.	Sample No.	Conclusion
D.O.B.			Tested by Date

Unless otherwise indicated, all cells are positive for Kp^b and Lu^b and negative for Wr^a and Co^b.

Please note some cellstab bottles are labelled R142 4401

Instructions for use can be found at <http://www.blood.co.uk/reagents>

	Rh	C	D	E	c	e	C ^w	M	N	S	s	P1	Lu ^a	K	k	Kp ^a	Le ^a	Le ^b	Fy ^a	Fy ^b	Jk ^a	Jk ^b	Other																
1	R ₁ ^w R ₁	+	+	0	0	+	+	+	0	+	0	1	0	0	+	0	0	+	0	+	+	0																	
2	R ₁ R ₁	+	+	0	0	+	0	0	+	0	+	4	0	+	+	0	0	+	+	0	0	+																	
3	R ₂ R ₂	0	+	+	+	0	0	+	0	+	0	2	0	0	+	0	0	0	0	+	0	+																	
4	r'r	+	0	0	+	+	0	+	0	0	+	0	0	0	+	0	+	0	0	+	+	0																	
5	r''r	0	0	+	+	+	0	0	+	0	+	0	0	0	+	0	0	+	+	0	+	0																	
6	rr	0	0	0	+	+	0	+	0	+	0	1	0	+	0	0	0	+	+	0	+	0																	
7	rr	0	0	0	+	+	0	+	0	0	+	1	+	+	+	0	+	0	0	+	+	0																	
8	rr	0	0	0	+	+	0	+	0	+	0	0	0	0	+	+	0	0	0	+	0	+																	
9	rr	0	0	0	+	+	0	0	+	+	0	0	0	0	+	0	+	0	0	+	0	+																	
10	rr	0	0	0	+	+	0	0	+	0	+	4	0	0	+	0	+	0	+	0	0	+	HLA+																
Reagent																																							
Lot No.																																							